EU Type Examination Certificate

This is to certify that:

Zagor Is Güv. Ekip. San. Tic. Ltd. Sti
Fevzi Cakmak Mahallesi
Mimar Sinan Caddesi
No:40 Dembeller Plaza
Istanbul, Pendik
34899
Turkey

Holds Certificate Number: CE 732023

In respect of:

Model ZGR 5030 and ZGR 5035 Face mask
To technical specification Annex II (EHSR) of the PPE Regulation (EU) 2016/425
PPE for use by healthcare professionals as per Commission recommendation 2020/403

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Drs. Dave Hagenaars, Managing Director

First Issued: 2020-11-12
Latest Issue: 2020-11-12

Effective Date: 2020-11-12
Expiry Date: 2021-11-12

PRODUCTS
RvA C 646

...making excellence a habit.™

EU Type Examination Certificate

No. CE 732023

Product Specification

Product Name:	Particulate Respirator.
Product Type:	Particulate filtering half masks for use by Healthcare professionals.
Model:	**ZGR 5030 and ZGR 5035**
Classification:	ZGR 5030 FFP3 NR valved ZGR 5035 FFP3 NR un-valved
Technical Specification:	Technical specification to satisfy Annex II of the PPE Regulation (EU) 2016/425.
Product Description:	The ZGR 5030 respirator is non-reusable, secured to the face of the user by a pair of adjustable ear straps, and has an exhalation valve. The ZGR 5035 respirator is non-reusable, secured to the face of the user by a pair of adjustable ear straps, and has no exhalation valve The respirators are FFP3 class, conical type. The respirator listed on this certificate is for use by healthcare workers, first responders and other personnel involved in the efforts to contain the COVID-19 virus and avoid its further spread. The product covered by this certificate is not approved for industrial applications and the certificate is only valid as long as EU Commission recommendation sheet 2020/403 remains applicable.
Product Assessments:	BSI's PPE for Healthcare Professionals 2020/403 – RPE Technical Specification.

First Issued: 2020-11-12
Latest Issue: 2020-11-12

Effective Date: 2020-11-12
Expiry Date: 2021-11-12

EU Type Examination Certificate

No. CE 732023

Certificate Administration Details

Technical File Reference: Zagor TF1 revision 4

Certificate Amendment Record:

Issue date	Comments	BSI Review No.
November 2020	First issue, rev. 01	2797:20:3253602

Certificate validity

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall process utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The validity of the Certificate for the products is also dependent on the maintenance of the EU Conformity to Type based on Internal Production Control plus supervised product checks at random intervals, Annex VII (Module C2) as referenced on BSI issued Certificate CE 732024.

First Issued: 2020-11-12
Latest Issue: 2020-11-12

Effective Date: 2020-11-12
Expiry Date: 2021-11-12